ENVIRONMENTAL HEALTH AND SAFETY
University of Nebraska-Lincoln

AUTOCLAVE LOG SHEET

Loads containing biohazardous waste must be autoclaved at ≥ 121°C with a sterlization time of at least 30 minutes.

Autoclave make/model:		**Location (Building):**	
Person Responsible for autoclave:		**Location (Room Number):**	
Contact Phone Number:		**EHS Autoclave ID:**	

Date	Load Contents	Cycle number	Cycle Type	Sterilization Time (min)	Pressure (psi)	Max Temp Reached	Tape Result (pass/fail)	Chemical Indicator* Result (pass/fail)	Biological Indicator Used? (Y/N)	Operator	Comments

**Chemical indicator = autoclave tape or chemical integrator pack.*

Created 4/2020