

(Revised 1/19)

DISPOSING OF BIOHAZARDOUS MATERIALS, INCLUDING RECOMBINANT NUCLEIC ACIDS

Scope

This SOP applies to all work at UNL that is subject to the **UNL Biosafety Guidelines**. The content of this SOP is based on requirements established by the following standards:

- *NIH Guidelines for Research Involving Recombinant or Synthetic Nucleic Acid Molecules (NIH Guidelines)*, National Institutes of Health
- *Biosafety in Microbiological and Biomedical Laboratories (BMBL)*, Centers for Disease Control and National Institutes of Health
- *Bloodborne Pathogens Standard, 29 CFR 1910.1030*, Occupational Safety and Health Administration

For the purposes of this SOP, the term “biohazard” applies to:

- Recombinant and Synthetic Nucleic Acids (r/sNA) in all forms, natural and synthetic (e.g., DNA, RNA, shRNA, etc.)
- Any waste material that is potentially infectious to humans, plants, or animals
- Genetically-modified organisms, including animals, plants, microorganisms and arthropods

General guidance as well as classes of biohazardous materials and their associated appropriate methods of disposal are described below.

General Guidance

Biohazardous waste that is autoclaved on-site to render it non-infectious or non-viable must be collected in bags that are rated for autoclave use and that have the universal biohazard symbol visible on the bag. Autoclave tape must be affixed to the container prior to treatment (select a lead-free autoclave tape); the tape should be placed over the biohazard symbol in a “X” pattern. Any biohazard labels and markings not covered by tape must be obliterated following treatment and prior to disposal. Autoclaved waste

(Created 3/03; Revised 6/07, 5/08, 12/11)

bags must be placed in unmarked outer bags that are dark or opaque (i.e., not see-through) prior to disposal, and all liquids must be absorbed onto a solid media prior to placement in refuse containers or if allowed, they may be disposed via the sanitary sewer. Sewer disposal is limited to nutrient media and dilute disinfectants that are otherwise non-hazardous (do not contain chemical constituents regulated under hazardous waste laws, such as heavy metals, solvents, etc.). Contact EHS if you are unsure if a liquid can be sewer disposed.

- Waste items that are chemically disinfected can be disposed of in the regular trash after sufficient contact time with the disinfectant. See the EHS SOP, ***Chemical Disinfectants for Biohazardous Materials*** for guidance on contact times.
- Biohazardous waste that is managed through EHS for off-site treatment must be accumulated in specific containers. Contact EHS for additional information.
- Autoclaves used to decontaminate biohazardous waste must be part of the EHS Autoclave Testing Program or users must conduct independent verification testing using biological indicators and log the testing. See the EHS SOP, ***Autoclave Operation and Performance Testing*** for guidance.



Do not use red biohazard bags for any type of waste collection other than biohazardous waste (i.e., regular trash, radioactive materials not contaminated with biohazards).

Specific Waste Types

- ***Human blood and other potentially infectious body fluids*** and contaminated materials must be decontaminated prior to disposal as solid waste refuse or managed through EHS. Generally, this type of waste is decontaminated on-site by autoclaving or collected by EHS and disposed via a vendor. This material must not be incinerated on-site.
- ***Laboratory waste from infectious agents*** (i.e., culture plate and broths) must be autoclaved prior to disposal. Chemical disinfection is generally suitable for contaminated disposable and non-disposable laboratory equipment and consumables (i.e., pipette tips, glass, cuvettes, etc.). This material must not be incinerated on-site.
- ***Recombinant nucleic acid-containing materials*** (i.e. cultures, microbes, plasmids, plants) must be decontaminated or inactivated before disposal. Chemical treatment with an appropriate disinfectant is an acceptable method of inactivation. Autoclaving of rNA-containing materials is also acceptable.
- ***Research animal carcasses*** and parts, contaminated bedding and feeds, excreta, etc. must be disposed via incineration or managed through EHS. Cages

and other non-removable, non-porous surfaces can be treated by chemical disinfection or autoclaved.



Note: For additional restrictions related to on-site incineration, refer to the EHS SOP, *Pathological Waste Incinerators - Operating Permit Requirements*.

- **Soils and other plant-related materials** Soil and plant material (i.e. leaves, shoots, stalks, etc.) that contains rNA or is contaminated with plant pathogens must be biologically inactivated prior to disposal. Inactivation by autoclaving requires sufficient run time and specific temperatures and pressures. Soil is more difficult to inactivate, as it is very dense. See the EHS SOP, **Autoclave Operation and Performance Testing** for guidance on autoclave parameters for inactivating soil and plant material.
- **Animal diagnostic specimen shipping containers**, if contaminated, must be treated by autoclaving or chemical treatment prior to disposal, incinerated, or managed through EHS for disposal.
- **Sharps**, including glass, razor blades, needles, and other objects that can penetrate the skin and are potentially contaminated with bloodborne pathogens or other biohazardous materials must be accumulated in rigid, leak-proof, labeled, containers specifically designed for use with bloodborne pathogens and purchased from a commercial supplier (i.e., red sharps boxes). Full containers



Figure 1 Contaminated Sharps Collection Containers

must be autoclaved or managed through EHS.

Use of red biohazard sharps boxes for collection of non-biohazardous sharps is discouraged. Green plastic sharp containers are commercially available for this purpose (see Figure 2). If a red biohazard plastic container is used for non-biohazardous sharps, the container must have the biohazard symbol thoroughly



Figure 2 Non-Infectious Sharps Containers
defaced and be clearly marked as “non-infectious.”

- **Pipette and Pipette Tips** contaminated with bloodborne pathogens or other biohazardous materials must be decontaminated prior to disposal. Decontamination can be achieved by autoclaving or chemical treatment with an appropriate disinfectant.
 - Contaminated pipettes and pipette tips to be decontaminated by autoclaving should be collected in rigid, leak-proof, sealable and labeled containers. The universal biohazard symbol should be visible on the container. See example containers in Figure 3A.
 - A sturdy cardboard box lined with a biohazard bag is an acceptable alternative collection container for these types of potential sharps (Figure 3B).
 - A shallow pan with a lid (Figure 3C) can be filled with a chemical disinfectant and used to collect contaminated pipettes and tips.
 - If collected after decontamination, the container should be rigid and sealable and not have the biohazard symbol visible on the outside of the container.

- A cardboard box lined with a trash bag is acceptable for collection of decontaminated pipettes



Figure 3 Contaminated Pipette and Pipette Tip collection containers